

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06182026-015

<p>Gage ID PEA-2233000331 Gage S/N 2233000331 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 6/18/2026 Next Due 12/18/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Calibrated black light serial no. 2234240 was also used during this calibration.
 Calibration due date for black light is on 06/15/2027

Black light is traceable to NIST by the following numbers: 272637, 272638, 272639, 272640

Standard ID	12"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6722	Nominal	7075	Maximum	7428
Ref Type	Long Wave	As Found	7248	Accuracy	173	Fail Before	No
		As Left	7248	Accuracy	173	Fail After	No
Gage ID of Standard	REF-1944278						
Std Due Date	6/15/2027	Gage S/N	1944278/1944280	NIST No.	272637, 272638, 272639, 272640		

Standard ID	15"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	4520	Nominal	4757	Maximum	4994
Ref Type	Long Wave	As Found	4925	Accuracy	168	Fail Before	No
		As Left	4925	Accuracy	168	Fail After	No
Gage ID of Standard	REF-1944278						
Std Due Date	6/15/2027	Gage S/N	1944278/1944280	NIST No.	272637, 272638, 272639, 272640		

Calibrated By Taichi Daimo

Signature 

Date: 6/18/2026

APPROVED
By Kayla Myers at 11:56 am, Jun 19, 2026