TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06182026-009

Gage ID	PEA-1902251622
Gage S/N	1902251622
Description	12" Digital Depth Gauge
Operating Procedure:	OP-123
Unit of Meas.	Inch
Manufacturer	Insize
Cal. Date	6/18/2026
Next Due	12/18/2026
Cal. Freq.	6.00 Months
Location	Lab

Environmental Conditions

Temperature	68 +/- 2 deg F
Humidity	20-55%

Approved	Yes
Customer Info.	Peak NDT

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	1.000 "	**Uncertainty**		**Units**		**Type**	V
Limited Use?	No	**Minimum**	0.999	**Nominal**	1.000	**Maximum**	1.001
Ref Type	GAGE BLOCKS	**As Found**	1.000	**Accuracy**	0.000	**Fail Before**	No
		As Left	1.000	**Accuracy**	0.000	**Fail After**	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	**Gage S/N**	061259	**NIST No.**	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Standard ID	3.000 "	**Uncertainty**		**Units**		**Type**	V
Limited Use?	No	**Minimum**	2.999	**Nominal**	3.000	**Maximum**	3.001
Ref Type	GAGE BLOCKS	**As Found**	3.000	**Accuracy**	0.000	**Fail Before**	No
		As Left	3.000	**Accuracy**	0.000	**Fail After**	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	**Gage S/N**	061259	**NIST No.**	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Standard ID	5.000 "	**Uncertainty**		**Units**		**Type**	V
Limited Use?	No	**Minimum**	4.999	**Nominal**	5.000	**Maximum**	5.001

Ref Type	GAGE BLOCKS	**As Found**	5.000	**Accuracy**	0.000	**Fail Before**	No
		As Left	5.000	**Accuracy**	0.000	**Fail After**	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	**Gage S/N**	061259	**NIST No.**	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Standard ID	7.000 "	**Uncertainty**		**Units**		**Type**	V
Limited Use?	No	**Minimum**	6.999	**Nominal**	7.000	**Maximum**	7.001
Ref Type	GAGE BLOCKS	**As Found**	7.000	**Accuracy**	0.000	**Fail Before**	No
		As Left	7.000	**Accuracy**	0.000	**Fail After**	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	**Gage S/N**	061259	**NIST No.**	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Standard ID	9.000 "	**Uncertainty**		**Units**		**Type**	V
Limited Use?	No	**Minimum**	8.999	**Nominal**	9.000	**Maximum**	9.001
Ref Type	GAGE BLOCKS	**As Found**	9.001	**Accuracy**	0.001	**Fail Before**	No
		As Left	9.001	**Accuracy**	0.001	**Fail After**	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	**Gage S/N**	061259	**NIST No.**	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Calibrated By **Taichi Daimo** **Signature** **Date:** **6/18/2026**

APPROVED

By Kayla Myers at 11:45 am, Jun 19, 2026