



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abs calibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83101
Mfg / Make	Insize	Model	1108-300
Serial Number	PEA-3006211617	Range	0-12"

Description	Digital Caliper
--------------------	-----------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
OD					
3.000	+/- .001	3.000	PASS	3.000	PASS
6.000	+/- .001	6.000	PASS	6.000	PASS
9.000	+/- .001	9.000	PASS	9.000	PASS
12.000	+/- .001	12.000	PASS	12.000	PASS
ID					
3.000	+/- .001	3.000	PASS	3.000	PASS
6.000	+/- .001	6.000	PASS	6.000	PASS
9.000	+/- .001	9.000	PASS	9.000	PASS
12.000	+/- .001	12.000	PASS	12.000	PASS
Depth					
3.000	+/- .001	3.000	PASS	3.000	PASS
6.000	+/- .001	6.000	PASS	6.000	PASS
9.000	+/- .001	9.000	PASS	9.000	PASS
12.000	+/- .001	12.000	PASS	12.000	PASS

Temperature	69° F	Humidity	49%
Calibration Date	07/14/2026	Calibration Due	01/14/2027

Technician	OM	Acceptance Criteria	See Above
-------------------	----	----------------------------	-----------

Description	Serial Number	Calibration Due Date
Master Plate	ABS-029	12/24/2028
Master Blocks	ABS-132	12/03/2027
Check Master	ABS-014	11/25/2029

The calibration measurements are performed in accordance with WI-8.01 Rev. 3 and Z540.-3-2006.
All measurements above were recorded at the time of calibration.
All instruments are traceable to NIST. ISO 9001:2105 Certified
This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager