



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abscalibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83105
Mfg / Make	R.B. Annis	Model	D-1558464376-1
Serial Number	PEA-11955	Range	-20 to +20g

Description	Gauss Meter
--------------------	-------------

Nominal g	Tolerance g	As Found g	Condition	As Left g	Condition
-20	+/- 1	-20	PASS	-20	PASS
-10	+/- 1	-10	PASS	-10	PASS
0	+/- 1	0	PASS	0	PASS
+10	+/- 1	+10	PASS	+10	PASS
+20	+/- 1	+20	PASS	+20	PASS

Temperature	69° F	Humidity	49%
Calibration Date	07/14/2026	Calibration Due	01/14/2027

Technician	OM	Acceptance Criteria	See Above
-------------------	----	----------------------------	-----------

Description	Serial Number	Calibration Due Date
Master Gauss Calibrator	ABS-142	01/28/2028
Master Gauss Calibrator	ABS-340	01/28/2028

The calibration measurements are performed in accordance with WI-8.39 Rev. 1 and Z540.-3-2006.
All measurements above were recorded at the time of calibration.
All instruments are traceable to NIST. ISO 9001:2105 Certified
This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan
Billy Quinlan
Lab Manager