



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abscalibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83109
Mfg / Make	Gagemaker	Model	M80-0403
Serial Number	PEA-RUN045	Range	-.010 to +.010"

Description	Lead Gauge
-------------	------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
-.010	+/- .0005	-.010	PASS	-.010	PASS
-.005	+/- .0005	-.005	PASS	-.005	PASS
.000	+/- .0005	.000	PASS	.000	PASS
+.005	+/- .0005	+.005	PASS	+.005	PASS
+.010	+/- .0005	+.010	PASS	+.010	PASS

Temperature	69° F	Humidity	49%
Calibration Date	07/14/2026	Calibration Due	07/14/2027

Technician	OM	Acceptance Criteria	See Above
------------	----	---------------------	-----------

Description	Serial Number	Calibration Due Date
Master Height	ABS-282	06/17/2029
Master Plate	ABS-029	12/24/2028

The calibration measurements are performed in accordance with WI-8.02 Rev. 2 and Z540.-3-2006.
All measurements above were recorded at the time of calibration.
All instruments are traceable to NIST. ISO 9001:2105 Certified
This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan
Billy Quinlan
Lab Manager

APPROVED
By Kayla Myers at 5:31 pm, Jul 16, 2026