



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abscalibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83108
Mfg / Make	Gagemaker	Model	LS-1007
Serial Number	PEA-JU21JL0077	Range	2 TPF, Even Pitch

Description	Lead Standard
-------------	---------------

Feature	Nominal "	Tolerance	As Found "	Condition	As Left "	Condition
Angle	30°	+/- 1°	30°	PASS	30°	PASS
Angle	30°	+/- 1°	30°	PASS	30°	PASS
Dim 1	.50173	+/- .0001	.50175	PASS	.50175	PASS
Dim 2	1.0034	+/- .0001	1.00345	PASS	1.00345	PASS
Dim 3	2.0069	+/- .0002	2.00695	PASS	2.00695	PASS
Dim 4	3.0104	+/- .0002	3.01045	PASS	3.01045	PASS
Dim 5	4.0138	+/- .0002	4.01385	PASS	4.01385	PASS
Radius	.030	+/- .001	.030	PASS	.030	PASS

Temperature	69° F	Humidity	49%
Calibration Date	07/14/2026	Calibration Due	07/14/2027
Technician	OM	Acceptance Criteria	See Above
Description	Serial Number	Calibration Due Date	
Master Comparator	ABS-362	05/16/2027	

The calibration measurements are performed in accordance with WI-8.38 Rev. 1 and Z540.-3-2006.

All measurements above were recorded at the time of calibration.

All instruments are traceable to NIST. ISO 9001:2015 Certified.

This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager

APPROVED

By Kayla Myers at 5:30 pm, Jul 16, 2026