



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abs calibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83106
Mfg / Make	Generic	Model	1018 Steel
Serial Number	PEA-J02108	Range	.100 - .500"

Description	Step Block
-------------	------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
.1000	+/- .0005	.100447	PASS	.100447	PASS
.2000	+/- .0005	.200354	PASS	.200354	PASS
.3000	+/- .0005	.300323	PASS	.300323	PASS
.4000	+/- .0005	.400489	PASS	.400489	PASS
.5000	+/- .0005	.500451	PASS	.500451	PASS

Temperature	69° F	Humidity	49%
Calibration Date	07/14/2026	Calibration Due	07/14/2027

Technician	OM	Acceptance Criteria	See Above
------------	----	---------------------	-----------

Description	Serial Number	Calibration Due Date
Master LH600	ABS-327	12/26/2027
Master Plate	ABS-350	12/05/2026

The calibration measurements are performed in accordance with WI-8.28 Rev. 2 and Z540.-3-2006.
All measurements above were recorded at the time of calibration.
All instruments are traceable to NIST. ISO 9001:2105 Certified
This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan
Billy Quinlan
Lab Manager

APPROVED
By Kayla Myers at 5:29 pm, Jul 16, 2026