



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@absclibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83134
Mfg / Make	Dakota	Model	ZX-3
Serial Number	PEA-23863	Range	0-6"

Description	UT Thickness Gauge
--------------------	--------------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
.100	+/- .005	.098	PASS	.098	PASS
.200	+/- .005	.199	PASS	.199	PASS
.300	+/- .005	.300	PASS	.300	PASS
.400	+/- .005	.400	PASS	.400	PASS
.500	+/- .005	.501	PASS	.501	PASS

Temperature	69° F	Humidity	49%
Calibration Date	07/14/2026	Calibration Due	01/14/2027

Technician	OM	Acceptance Criteria	See Above
-------------------	----	----------------------------	-----------

Description	Serial Number	Calibration Due Date
Master Block	ABS-095	11/25/2027
Master Plate	ABS-029	12/24/2028

The calibration measurements are performed in accordance with WI-8.33 Rev. 2 and Z540.-3-2006.
All measurements above were recorded at the time of calibration.
All instruments are traceable to NIST. ISO 9001:2105 Certified
This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager