



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abscalibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83103
Mfg / Make	Generic	Model	DLM200
Serial Number	PEA-240302844	Range	0-20,000 fc

Description	White Light Meter
--------------------	-------------------

Distance	Nominal fc	Tolerance	As Found fc	Condition	As Left fc	Condition
White						
12"	703	+/-10%	709	PASS	709	PASS
18"	387	+/-10%	418	PASS	418	PASS
24"	206	+/-10%	211	PASS	211	PASS

Temperature	69° F	Humidity	49%
Calibration Date	07/14/2026	Calibration Due	01/14/2027

Technician	OM	Acceptance Criteria	See Above
-------------------	----	----------------------------	-----------

Description	Serial Number	Calibration Due Date
Master Light Meter	ABS-285	05/05/2027
Master Light Meter	ABS-274	09/18/2027
Master Plate	ABS-029	12/24/2028

The calibration measurements are performed in accordance with WI-8.34 Rev. 2 and Z540.-3-2006.

All measurements above were recorded at the time of calibration.

All instruments are traceable to NIST. ISO 9001:2015 Certified.

This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager