

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-03042026-019

| | |
|---|--|
| <p>Gage ID PEA-0603253822 Gage S/N 0603253822 Description 12" Digital Caliper Operating Procedure: OP-110 Unit of Meas. inch Manufacturer Insize Cal. Date 3/4/2026 Next Due 9/4/2026 Cal. Freq. 6.00 Months Location Lab</p> | <p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p> |
|---|--|

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

| | | | |
|-----------------------------|-----------------------|-----------------------|-----------------------|
| Standard ID 1.000 " | Uncertainty | Units | Type V |
| Limited Use? No | Minimum 0.999 | Nominal 1.000 | Maximum 1.001 |
| Ref Type GAGE BLOCKS | As Found 1.000 | Accuracy 0.000 | Fail Before No |
| | As Left 1.000 | Accuracy 0.000 | Fail After No |

Gage ID of Standard REF-GBS-061259
Std Due Date 12/3/2027 **Gage S/N** 061259 **NIST No.** LMUA-131223 2018-3-20, AFS-276 2018-3-20

| | | | |
|-----------------------------|-----------------------|-----------------------|-----------------------|
| Standard ID 3.000 " | Uncertainty | Units | Type V |
| Limited Use? No | Minimum 2.999 | Nominal 3.000 | Maximum 3.001 |
| Ref Type GAGE BLOCKS | As Found 3.000 | Accuracy 0.000 | Fail Before No |
| | As Left 3.000 | Accuracy 0.000 | Fail After No |

Gage ID of Standard REF-GBS-061259
Std Due Date 12/3/2027 **Gage S/N** 061259 **NIST No.** LMUA-131223 2018-3-20, AFS-276 2018-3-20

| | | | |
|----------------------------|----------------------|----------------------|----------------------|
| Standard ID 5.000 " | Uncertainty | Units | Type V |
| Limited Use? No | Minimum 4.999 | Nominal 5.000 | Maximum 5.001 |

| | | | | | | | |
|---------------------|----------------|----------|--------|----------|--|-------------|----|
| Ref Type | GAGE BLOCKS | As Found | 5.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 5.000 | Accuracy | 0.000 | Fail After | No |
| Gage ID of Standard | REF-GBS-061259 | | | | | | |
| Std Due Date | 12/3/2027 | Gage S/N | 061259 | NIST No. | LMUA-131223 2018-3-20, AFS-276 2018-3-20 | | |

| | | | | | | | |
|--------------|-------------|-------------|-------|----------|-------|-------------|-------|
| Standard ID | 7.000 " | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 6.999 | Nominal | 7.000 | Maximum | 7.001 |
| Ref Type | GAGE BLOCKS | As Found | 7.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 7.000 | Accuracy | 0.000 | Fail After | No |

| | | | | | | | |
|---------------------|----------------|----------|--------|----------|--|--|--|
| Gage ID of Standard | REF-GBS-061259 | | | | | | |
| Std Due Date | 12/3/2027 | Gage S/N | 061259 | NIST No. | LMUA-131223 2018-3-20, AFS-276 2018-3-20 | | |

| | | | | | | | |
|--------------|-------------|-------------|-------|----------|-------|-------------|-------|
| Standard ID | 9.000 " | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 8.999 | Nominal | 9.000 | Maximum | 9.001 |
| Ref Type | GAGE BLOCKS | As Found | 9.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 9.000 | Accuracy | 0.000 | Fail After | No |

| | | | | | | | |
|---------------------|----------------|----------|--------|----------|--|--|--|
| Gage ID of Standard | REF-GBS-061259 | | | | | | |
| Std Due Date | 12/3/2027 | Gage S/N | 061259 | NIST No. | LMUA-131223 2018-3-20, AFS-276 2018-3-20 | | |

Calibrated By Taichi Daimo Signature  Date: 3/4/2026

APPROVED
By Kayla Myers at 1:37 pm, Mar 09, 2026