

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-03062026-014

<p>Gage ID PEA-2509255017 Gage S/N 2509255017 Description 12" Digital Depth Gauge Operating Procedure: OP-123 Unit of Meas. inch Manufacturer Insize Cal. Date 3/6/2026 Next Due 9/6/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID 1.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.999	Nominal 1.000	Maximum 1.001
Ref Type GAGE BLOCKS	As Found 1.000	Accuracy 0.000	Fail Before No
	As Left 1.000	Accuracy 0.000	Fail After No
Gage ID of Standard REF-GBS-061259	Gage S/N 061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20
Std Due Date 12/3/2027			

Standard ID 3.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 2.999	Nominal 3.000	Maximum 3.001
Ref Type GAGE BLOCKS	As Found 3.000	Accuracy 0.000	Fail Before No
	As Left 3.000	Accuracy 0.000	Fail After No
Gage ID of Standard REF-GBS-061259	Gage S/N 061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20
Std Due Date 12/3/2027			

Standard ID 5.000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 4.999	Nominal 5.000	Maximum 5.001

Ref Type	GAGE BLOCKS	As Found	5.000	Accuracy	0.000	Fail Before	No
		As Left	5.000	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	GAGE BLOCKS	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Standard ID	9.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	8.999	Nominal	9.000	Maximum	9.001
Ref Type	GAGE BLOCKS	As Found	9.000	Accuracy	0.000	Fail Before	No
		As Left	9.000	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Calibrated By Taichi Daimo Signature  Date: 3/6/2026

APPROVED
By Kayla Myers at 1:15 pm, Mar 09, 2026