

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-03042026-014

<p>Gage ID PEA-15652-1 Gage S/N 15652-1 Description 50-0-50 Gauss Meter Operating Procedure: OP-117 Unit of Meas. Gauss Manufacturer R. B. Annis Co., Inc. Cal. Date 3/4/2026 Next Due 9/4/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID +25	Uncertainty	Units	Type V
Limited Use? No	Minimum 24	Nominal 25	Maximum 26
Ref Type Gauss Meter	As Found 25	Accuracy 0	Fail Before No
	As Left 25	Accuracy 0	Fail After No
Gage ID of Standard REF-09-1234	Gage S/N 09-1234	NIST No. 1114538	
Std Due Date 5/12/2026			

Standard ID +50	Uncertainty	Units	Type V
Limited Use? No	Minimum 49	Nominal 50	Maximum 51
Ref Type Gauss Meter	As Found 50	Accuracy 0	Fail Before No
	As Left 50	Accuracy 0	Fail After No
Gage ID of Standard REF-09-1234	Gage S/N 09-1234	NIST No. 1114538	
Std Due Date 5/12/2026			

Standard ID -25	Uncertainty	Units	Type V
Limited Use? No	Minimum 24	Nominal 25	Maximum 26

Ref Type	Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No

Gage ID of Standard	REF-09-1234						
Std Due Date	5/12/2026	Gage S/N	09-1234	NIST No.	1114538		

Standard ID	-50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No

Gage ID of Standard	REF-09-1234						
Std Due Date	5/12/2026	Gage S/N	09-1234	NIST No.	1114538		

Calibrated By Taichi Daimo Signature  Date: 3/4/2026

APPROVED
By Kayla Myers at 12:51 pm, Mar 09, 2026