

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-03042026-015

Gage ID PEA-15652-2 Gage S/N 15652-2 Description 50-0-50 Gauss Meter Operating Procedure: OP-117 Unit of Meas. Gauss Manufacturer R. B. Annis Co., Inc. Cal. Date 3/4/2026 Next Due 9/4/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	+25	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	24	Nominal	25	Maximum	26
Ref Type	Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No
Gage ID of Standard	REF-09-1234						
Std Due Date	5/12/2026	Gage S/N	09-1234	NIST No.	1114538		
Standard ID	+50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No
Gage ID of Standard	REF-09-1234						
Std Due Date	5/12/2026	Gage S/N	09-1234	NIST No.	1114538		
Standard ID	-25	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	24	Nominal	25	Maximum	26

Ref Type	Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No

Gage ID of Standard REF-09-1234

Std Due Date 5/12/2026

Gage S/N 09-1234

NIST No. 1114538

Standard ID	-50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No

Gage ID of Standard REF-09-1234

Std Due Date 5/12/2026

Gage S/N 09-1234

NIST No. 1114538

Calibrated By Taichi Daimo

Signature



Date:

3/4/2026

APPROVED*By Kayla Myers at 12:52 pm, Mar 09, 2026*