

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-01122026-016

<b>Gage ID</b> PEA-RUN045 <b>Gage S/N</b> RUN045 <b>Description</b> Lead Gauge <b>Operating Procedure:</b> OP-124 <b>Unit of Meas.</b> Inch <b>Manufacturer</b> Gagemaker <b>Cal. Date</b> 1/12/2026 <b>Next Due</b> 7/12/2026 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> Temperature 68 +/- 2 deg F Humidity 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
---	---

#### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### Findings

<b>Standard ID</b>	+0.00500"	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	Minimum	0.00450	Nominal	0.00500	Maximum	0.00550
<b>Ref Type</b>	Mic Trac	As Found	0.00492	Accuracy	-0.00008	Fail Before	No
		As Left	0.00492	Accuracy	-0.00008	Fail After	No
<b>Gage ID of Standard</b>	REF-G141308						
<b>Std Due Date</b>	10/1/2026	<b>Gage S/N</b>	G141308	<b>NIST No.</b>	747V66-240502-00		
<b>Standard ID</b>	+0.01000"	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	Minimum	0.00950	Nominal	0.01000	Maximum	0.01050
<b>Ref Type</b>	Mic Trac	As Found	0.00993	Accuracy	-0.00007	Fail Before	No
		As Left	0.00993	Accuracy	-0.00007	Fail After	No
<b>Gage ID of Standard</b>	REF-G141308						
<b>Std Due Date</b>	10/1/2026	<b>Gage S/N</b>	G141308	<b>NIST No.</b>	747V66-240502-00		
<b>Standard ID</b>	-0.00500"	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	Minimum	-0.00550	Nominal	-0.00500	Maximum	-0.00450

Ref Type	Mic Trac	As Found	-0.00506	Accuracy	-0.00006	Fail Before	No
		As Left	-0.00506	Accuracy	-0.00006	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 10/1/2026 Gage S/N G141308 NIST No. 747V66-240502-00

Standard ID	-0.01000"	Uncertainty	Units	Type	V		
Limited Use?	No	Minimum	-0.01050	Nominal	-0.01000	Maximum	-0.00950
Ref Type	Mic Trac	As Found	-0.01003	Accuracy	-0.00003	Fail Before	No
		As Left	-0.01003	Accuracy	-0.00003	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 10/1/2026 Gage S/N G141308 NIST No. 747V66-240502-00

---

Calibrated By Taichi Daimo Signature  Date: 1/12/2026