



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abs calibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	84007
Mfg / Make	Mitutoyo	Model	CD-24" C
Serial Number	PEA-0021721	Range	0-24"

Description	Digital Caliper
-------------	-----------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
OD					
6.000	+/- .002	6.000	PASS	6.000	PASS
12.000	+/- .002	12.000	PASS	12.000	PASS
18.000	+/- .002	18.000	PASS	18.000	PASS
24.000	+/- .002	24.000	PASS	24.000	PASS
ID					
6.000	+/- .002	6.000	PASS	6.000	PASS
12.000	+/- .002	12.000	PASS	12.000	PASS
18.000	+/- .002	18.000	PASS	18.000	PASS
24.000	+/- .002	24.000	PASS	24.000	PASS

Temperature	69° F	Humidity	49%
Calibration Date	08/11/2026	Calibration Due	02/11/2027

Technician	AR	Acceptance Criteria	See Above
------------	----	---------------------	-----------

Description	Serial Number	Calibration Due Date
Master Block	ABS-132	12/03/2027
Master Plate	ABS-350	12/05/2026
Check Master	ABS-356	05/01/2030

The calibration measurements are performed in accordance with WI-8.01 Rev. 3 and Z540.-3-2006.

All measurements above were recorded at the time of calibration.

All instruments are traceable to NIST. ISO 9001:2105 Certified

This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager

APPROVED

By Kayla Myers at 11:10 am, Sep 17, 2026