



Tubular Inspection Products

1551 Windway St.
Odessa, TX 79763
(432) 363-2200

PO Box 842
Odessa, TX 79760
Fax: (432) 363-2202

CERTIFICATE OF CALIBRATION

Traceable to NIST — ISO/IEC 17025:2017

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO 17025:2017 • ISO 9001 • API Q1/7-2

INSTRUMENT & CALIBRATION DETAILS

Customer	Peak NDT Solutions, LLC	Manufacturer	Gagemaker
Calibration Date	August 01, 2026	TIP Procedure	TIP-K80PG
Calibration Due Date	August 01, 2027	Calibration Technician	Ryan Britton
Serial Number	JZ29JF0009	Order Number	81126
Instrument	PD-3004 Pit Gauge	Temperature / Humidity	68.0°F / RH 40.0%

CALIBRATION RESULTS

Reference Size	Actual Reading	Deviation	Tolerance	Result
0.100"	0.099"	0.001"	±0.002"	PASS
0.200"	0.199"	0.001"	±0.002"	PASS
0.300"	0.299"	0.001"	±0.002"	PASS

INSTRUMENT STATUS SUMMARY

Instrument Received	Instrument Returned	Instrument Status
In Tolerance	In Tolerance	PASSED

CALIBRATION STANDARDS USED (TRACEABILITY)

- TIP Master Optical Comparator, Keyence IM-8030, S/N 4C0Q000412 (Cert. Due 01/28/2027)
- TIP Master Hygrometer, S/N 31874 / 1H2048MCR / 130070751 (Cert. Due 01/28/2027)

Calibration Technician

Ryan Britton

Ryan Britton

Issue Date: August 01, 2026

APPROVED

By Kayla Myers at 2:56 pm, Sep 03, 2026