

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-01272025-010

Gage ID PEA-12941-4 Gage S/N 12941-4 Description 50-0-50 Gauss Meter Operating Procedure: OP-117 Unit of Meas. Gauss Manufacturer R. B. Annis Co., Inc. Cal. Date 1/27/2025 Next Due 7/27/2025 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% <div style="text-align: center;"> Approved Yes Customer Info. Peak NDT </div>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	+25	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	24	Nominal	25	Maximum	26
Ref Type	Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No
Gage ID of Standard	REF-09-1234						
Std Due Date	5/10/2025	Gage S/N	09-1234	NIST No.	101912		

Standard ID	+50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No
Gage ID of Standard	REF-09-1234						
Std Due Date	5/10/2025	Gage S/N	09-1234	NIST No.	101912		

Standard ID	-25	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	24	Nominal	25	Maximum	26

Gage ID: PEA-12941-4

Page 2 of 2

Ref Type	Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No

Gage ID of Standard REF-09-1234

Std Due Date	5/10/2025	Gage S/N	09-1234	NIST No.	101912
--------------	-----------	----------	---------	----------	--------

Standard ID	-50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No

Gage ID of Standard REF-09-1234

Std Due Date	5/10/2025	Gage S/N	09-1234	NIST No.	101912
--------------	-----------	----------	---------	----------	--------

Calibrated By	Taichi Daimo	Signature		Date:	1/27/2025
---------------	--------------	-----------	--	-------	-----------

APPROVED

By Kayla Myers at 10:25 am, Jan 28, 2025