

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-01122026-012

<p><b>Gage ID</b> PEA-11955  <b>Gage S/N</b> 11955  <b>Description</b> 20-0-20 Gauss Meter  <b>Operating Procedure:</b> OP-117  <b>Unit of Meas.</b> Gauss  <b>Manufacturer</b> R. B. Annis Co., Inc.  <b>Cal. Date</b> 1/12/2026  <b>Next Due</b> 7/12/2026  <b>Cal. Freq.</b> 6.00                      Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT</p>
---	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

<b>Standard ID</b>	+10	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	9	<b>Nominal</b>	10	<b>Maximum</b>	11
<b>Ref Type</b>	Gauss Meter	<b>As Found</b>	10	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	10	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	Ref-09-1229						
<b>Std Due Date</b>	10/13/2026	<b>Gage S/N</b>	09-1229	<b>NIST No.</b>	114537		

<b>Standard ID</b>	+20	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	19	<b>Nominal</b>	20	<b>Maximum</b>	21
<b>Ref Type</b>	Gauss Meter	<b>As Found</b>	20	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	20	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	Ref-09-1229						
<b>Std Due Date</b>	10/13/2026	<b>Gage S/N</b>	09-1229	<b>NIST No.</b>	114537		

<b>Standard ID</b>	-10	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	9	<b>Nominal</b>	10	<b>Maximum</b>	11

Ref Type	Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/13/2026

Gage S/N 09-1229

NIST No. 114537

Standard ID	-20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/13/2026

Gage S/N 09-1229

NIST No. 114537

Calibrated By Taichi Daimo

Signature



Date: 1/12/2026

**APPROVED**

*By Kayla Myers at 3:03 pm, Jan 12, 2026*