Gage ID: PEA-11955 Page 1 of 2

# TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

## **Certificate of Calibration**

## Certificate No TSP-07312025-000

Gage ID PEA-11955

Gage S/N 11955

**Description** 20-0-20 Gauss Meter

Operating OP-117 Procedure:

Unit of Meas. Gauss

Manufacturer R. B. Annis Co., Inc.

Cal. Date 7/31/2025 Next Due 1/31/2026

Cal. Freq. 6.00

**Months** 

Location Lab

**Environmental Conditions** 

Temperature

68 +/- 2 deg F

Humidity

20-55%

**Approved** Yes Customer Info. Peak NDT

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### **Findings**

Standard ID	+10		Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	
Ref Type	Gauss Meter		As Found		10	Accuracy		0	Fail Before	No	
			As Left		10	Accuracy		0	Fail After	No	
Gage ID of Star	ndard	Ref-09-1229									
Std Due Date		10/10/2025	Gage S/N	09-1229	NIST No.		114537				
Standard ID	+20	<del></del>	Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		19	Nominal		20	Maximum	21	
Ref Type	Gauss	s Meter	As Found		20	Accuracy		0	Fail Before	No	
			As Left		20	Accuracy		0	Fail After	No	
Gage ID of Star	ndard	Ref-09-1229									
Std Due Date		10/10/2025	Gage S/N	09-1229	NIST No.		114537				
Standard ID	-10		Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	

Ref Type Gauss Meter As Found 10 Accuracy 0 Fail Before No As Left Accuracy Fail After No Gage ID of Standard Ref-09-1229 Std Due Date 10/10/2025 Gage S/N 09-1229 NIST No. 114537 Standard ID -20 Uncertainty Units Type V Limited Use? No Minimum 19 Nominal 20 Maximum 21 Ref Type Gauss Meter As Found 20 Accuracy 0 Fail Before No As Left Fail After No 20 Accuracy Gage ID of Standard Ref-09-1229 Std Due Date 10/10/2025 Gage S/N 09-1229 NIST No. 114537 Calibrated By Taichi Daimo Signature Date: 7/31/2025

Page

2 of 2

**APPROVED** 

Gage ID: PEA-11955

By Kayla Myers at 1:48 pm, Jul 31, 2025