

### TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

Certificate No TSP-01122026-013

<p><b>Gage ID</b> PEA-23863  <b>Gage S/N</b> 23863  <b>Description</b> UT Thickness Gage  <b>Operating Procedure:</b> OP-121  <b>Unit of Meas.</b> inch  <b>Manufacturer</b> Dakota Ultrasonics  <b>Cal. Date</b> 1/12/2026  <b>Next Due</b> 7/12/2026  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved Yes</b>  <b>Customer Info.</b> Peak NDT</p>
--	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

<b>Standard ID</b> 0.250"	<b>Uncertainty</b>	<b>Units</b>	<b>Type V</b>
<b>Limited Use?</b> No	<b>Minimum</b> 0.248	<b>Nominal</b> 0.250	<b>Maximum</b> 0.252
<b>Ref Type</b> 8 Step Block	<b>As Found</b> 0.250	<b>Accuracy</b> 0.000	<b>Fail Before</b> No
	<b>As Left</b> 0.250	<b>Accuracy</b> 0.000	<b>Fail After</b> No
<b>Gage ID of Standard</b> Ref-419719	<b>Gage S/N</b> 419719	<b>NIST No.</b> 176480	
<b>Std Due Date</b> 12/11/2026			

<b>Standard ID</b> 0.750"	<b>Uncertainty</b>	<b>Units</b>	<b>Type V</b>
<b>Limited Use?</b> No	<b>Minimum</b> 0.748	<b>Nominal</b> 0.750	<b>Maximum</b> 0.752
<b>Ref Type</b> 8 Step Block	<b>As Found</b> 0.752	<b>Accuracy</b> 0.002	<b>Fail Before</b> No
	<b>As Left</b> 0.752	<b>Accuracy</b> 0.002	<b>Fail After</b> No
<b>Gage ID of Standard</b> Ref-419719	<b>Gage S/N</b> 419719	<b>NIST No.</b> 176480	
<b>Std Due Date</b> 12/11/2026			

<b>Standard ID</b> 1.000 "	<b>Uncertainty</b>	<b>Units</b>	<b>Type V</b>
<b>Limited Use?</b> No	<b>Minimum</b> 0.998	<b>Nominal</b> 1.000	<b>Maximum</b> 1.002

Ref Type	8 Step Block	As Found	1.002	Accuracy	0.002	Fail Before	No
		As Left	1.002	Accuracy	0.002	Fail After	No

Gage ID of Standard Ref-419719  
 Std Due Date 12/11/2026

Gage S/N 419719 NIST No. 176480

Standard ID	1.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	1.498	Nominal	1.500	Maximum	1.502
Ref Type	8 Step Block	As Found	1.501	Accuracy	0.001	Fail Before	No
		As Left	1.501	Accuracy	0.001	Fail After	No

Gage ID of Standard Ref-419719  
 Std Due Date 12/11/2026

Gage S/N 419719 NIST No. 176480

Standard ID	2.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	1.998	Nominal	2.000	Maximum	2.002
Ref Type	8 Step Block	As Found	2.000	Accuracy	0.000	Fail Before	No
		As Left	2.000	Accuracy	0.000	Fail After	No

Gage ID of Standard Ref-419719  
 Std Due Date 12/11/2026

Gage S/N 419719 NIST No. 176480

Calibrated By Taichi Daimo Signature  Date: 1/12/2026

**APPROVED**  
 By Kayla Myers at 3:03 pm, Jan 12, 2026