

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration**Certificate No TSP-07312025-007**

Gage ID PEA-2209000313 Gage S/N 2209000313 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 7/31/2025 Next Due 1/31/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Calibrated black light serial no. 2234240 was also used during this calibration.
Calibration due date for black light is on 06/17/2026

Black light is traceable to NIST by the following numbers: 263546, 263547

Standard ID 12"	Uncertainty	Units	Type V
Limited Use? No	Minimum 6631	Nominal 6979	Maximum 7327
Ref Type Long Wave	As Found 7098	Accuracy 119	Fail Before No
	As Left 7098	Accuracy 119	Fail After No
Gage ID of Standard REF-1944278			
Std Due Date 6/17/2026	Gage S/N 1944278/1944 280	NIST No. 263546, 263547	

Standard ID 15"	Uncertainty	Units	Type V
Limited Use? No	Minimum 4468	Nominal 4703	Maximum 4938
Ref Type Long Wave	As Found 4824	Accuracy 121	Fail Before No
	As Left 4824	Accuracy 121	Fail After No
Gage ID of Standard REF-1944278			
Std Due Date 6/17/2026	Gage S/N 1944278/1944 280	NIST No. 263546, 263547	

Calibrated By Taichi Daimo Signature  Date: 7/31/2025

APPROVED
By Kayla Myers at 1:55 pm, Jul 31, 2025