



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@absclibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83465
Mfg / Make	RB Annis	Model	D-1558764376-1
Serial Number	PEA-41321	Range	-50 to +50g

Description	Gauss Meter
-------------	-------------

Nominal g	Tolerance g	As Found g	Condition	As Left g	Condition
+50	+/- .5	+60	FAIL	+50	PASS
+25	+/- .5	+35	FAIL	+25	PASS
0	+/- .5	+10	FAIL	0	PASS
-25	+/- .5	-15	FAIL	-25	PASS
-50	+/- .5	-40	FAIL	-50	PASS

****BOLD INDICATES AN OUT OF TOLERANCE CONDITION FOUND****

Temperature	69° F	Humidity	49%
Calibration Date	07/24/2026	Calibration Due	01/24/2027

Technician	CH	Acceptance Criteria	See Above
------------	----	---------------------	-----------

Description	Serial Number	Calibration Due Date
Master Meter	ABS-232	08/16/2026
Master Meter	ABS-340	01/28/2028

The calibration measurements are performed in accordance with WI-8.39 Rev.1 and Z540.-3-2006.

All measurements above were recorded at the time of calibration.

All instruments are traceable to NIST. ISO 9001:2105 Certified

This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager

APPROVED

By Kayla Myers at 10:15 am, Aug 07, 2026