



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abscalibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83467
Mfg / Make	Insize	Model	3287-1
Serial Number	PEA-220942452	Range	0-1"

Description	Groove Mic
-------------	------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
OD					
.250	+/- .001	.245	FAIL	.250	PASS
.500	+/- .001	.495	FAIL	.500	PASS
.750	+/- .001	.745	FAIL	.750	PASS
1.000	+/- .001	.995	FAIL	1.000	PASS
ID					
.250	+/- .001	.245	FAIL	.250	PASS
.500	+/- .001	.495	FAIL	.500	PASS
.750	+/- .001	.745	FAIL	.750	PASS
1.000	+/- .001	.995	FAIL	1.000	PASS

****BOLD INDICATES AN OUT OF TOLERANCE CONDITION FOUND****

Temperature	69° F	Humidity	49%
Calibration Date	07/24/2026	Calibration Due	01/24/2027

Technician	CH	Acceptance Criteria	See Above
------------	----	---------------------	-----------

Description	Serial Number	Calibration Due Date
Master Blocks	ABS-132	12/03/2027
Master Plate	ABS-350	12/05/2026
Master Mic	ABS-017	10/17/2026

The calibration measurements are performed in accordance with WI-8.05 Rev. 2 and Z540.-3-2006.

All measurements above were recorded at the time of calibration.

All instruments are traceable to NIST. ISO 9001:2105 Certified

This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager

APPROVED

By Kayla Myers at 10:11 am, Aug 07, 2026