

**TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

**Certificate of Calibration**

Certificate No TSP-06012023-024

<b>Gage ID</b> PEA-210700880 <b>Gage S/N</b> 210700880 <b>Description</b> Visible Light Meter <b>Operating Procedure:</b> OP-130 <b>Unit of Meas.</b> Foot Candles <b>Manufacturer</b> General <b>Cal. Date</b> 6/1/2023 <b>Next Due</b> 12/1/2023 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b>  Temperature 68 +/- 2 deg F Humidity 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------------------

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

SO 11793 JOB 12555

Using calibrated light meter serial no. 1944278 with sensor 1944280 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	95.3	96.9
Low	32.8	33.5

Readings As Left (foot candles)

	Master	Actual
High	95.3	96.9
Low	32.8	33.5

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater

Cal due date for light meter is 06/20/2023

NIST Traceable through : 210558, 210559, 210560

**APPROVED****By Kayla Myers at 2:29 pm, Jun 01, 2023**Calibrated By Taichi DaimoSignature Date: 6/1/2023