



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abscalibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83470
Mfg / Make	Gagemaker	Model	LG-6002
Serial Number	PEA-BR10LL0050	Range	-.010 - +.010

Description	Lead Gauge
-------------	------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
-.010	+/- .001	-.010	PASS	-.010	PASS
-.005	+/- .001	-.005	PASS	-.005	PASS
0	+/- .001	0	PASS	0	PASS
+.005	+/- .001	+.005	PASS	+.005	PASS
+.010	+/- .001	+.010	PASS	+.010	PASS

Temperature	69° F	Humidity	49%
Calibration Date	07/24/2026	Calibration Due	01/24/2027

Technician	CH	Acceptance Criteria	See Above
------------	----	---------------------	-----------

Description	Serial Number	Calibration Due Date
Master Plate	ABS-029	12/24/2028
Master Height	ABS-282	06/17/2029

The calibration measurements are performed in accordance with WI-8.02 Rev.2 and Z540.-3-2006.

All measurements above were recorded at the time of calibration.

All instruments are traceable to NIST. ISO 9001:2105 Certified

This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager

APPROVED

By Kayla Myers at 10:33 am, Aug 07, 2026