



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abs calibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83477
Mfg / Make	Insize	Model	4801-15E
Serial Number	PEA-2907248048	Range	1/32 – 1/4"

Description	Radius Gauge
--------------------	---------------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
Int / Ext		Int / Ext		Int / Ext	
.3125 / .3125	+/- .001	.3125 / .3125	PASS	.3125 / .3125	PASS
.46875 / .46875	+/- .001	.46875 / .46875	PASS	.46875 / .46875	PASS
.0625 / .0625	+/- .001	.0625 / .0625	PASS	.0625 / .0625	PASS
.07812 / .07812	+/- .001	.07812 / .07812	PASS	.07812 / .07812	PASS
.09375 / .09375	+/- .001	.09375 / .09375	PASS	.09375 / .09375	PASS
.109375 / .109375	+/- .001	.109375 / .109375	PASS	.109375 / .109375	PASS
0.1250 / .1250	+/- .001	0.1250 / .1250	PASS	0.1250 / .1250	PASS
.140625 / .140625	+/- .001	.140625 / .140625	PASS	.140625 / .140625	PASS
.15625 / .15625	+/- .001	.15625 / .15625	PASS	.15625 / .15625	PASS
.171875 / .171875	+/- .001	.171875 / .171875	PASS	.171875 / .171875	PASS
.1875 / .1875	+/- .001	.1875 / .1875	PASS	.1875 / .1875	PASS
.203125 / .203125	+/- .001	.203125 / .203125	PASS	.203125 / .203125	PASS
.21875 / .21875	+/- .001	.21875 / .21875	PASS	.21875 / .21875	PASS
.234375 / .234375	+/- .001	.234375 / .234375	PASS	.234375 / .234375	PASS
.250 / .250	+/- .001	.250 / .250	PASS	.250 / .250	PASS

Temperature	69° F	Humidity	50%
Calibration Date	07/24/2026	Calibration Due	07/24/2027

Technician	OM	Acceptance Criteria	See Above
-------------------	----	----------------------------	-----------

Description	Serial Number	Calibration Due Date
Master Comparator	ABS-362	05/16/2027

The calibration measurements are performed in accordance with WI-8.07 Rev. 2 and Z540.-3-2006.
All measurements above were recorded at the time of calibration.
All instruments are traceable to NIST. ISO 9001:2015 Certified.
This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager

APPROVED

By Kayla Myers at 10:29 am, Aug 07, 2026