



10386 Tanyard Rd
Willis, Tx. 77378
936-718-2047 / 936-537-7372
billy@abscalibration.com

CERTIFICATE OF CALIBRATION

Customer	Peak NDT	Cert Number	83471
Mfg / Make	Generic	Model	Working
Serial Number	PEA-E06373	Range	.100 - .500"

Description	Step Block
-------------	------------

Nominal "	Tolerance "	As Found "	Condition	As Left "	Condition
.1000	+/- .0005	.104060	FAIL	.100474	PASS
.2000	+/- .0005	.203326	FAIL	.200488	PASS
.3000	+/- .0005	.302206	FAIL	.300178	PASS
.4000	+/- .0005	.401354	FAIL	.400051	PASS
.5000	+/- .0005	.500179	PASS	.499762	PASS

****BOLD INDICATES AN OUT OF TOLERANCE CONDITION FOUND****

Temperature	69° F	Humidity	49%
Calibration Date	07/24/2026	Calibration Due	07/24/2027

Technician	OM	Acceptance Criteria	See Above
------------	----	---------------------	-----------

Description	Serial Number	Calibration Due Date
Master LH600	ABS-327	12/26/2027
Master Plate	ABS-350	12/05/2026

The calibration measurements are performed in accordance with WI-8.28 Rev. 2 and Z540.-3-2006.
All measurements above were recorded at the time of calibration.
All instruments are traceable to NIST. ISO 9001:2105 Certified
This report shall not be reproduced except in full, without the written consent of Absolute Calibration.

Billy Quinlan

Billy Quinlan
Lab Manager

APPROVED

By Kayla Myers at 10:37 am, Aug 07, 2026