
Company	Peak NDT Solutions	Doc #:	185536	Print#:	
Address	515A Verot School Road	OEM #:		Serial#:	2105201159
	Lafayette LA	Contact		Location	Calibration Lab
Zip:	70506	Dept	WO# 070726-11	Mfg.:	Insize
Page	1 of 1	Cal Date:	07/08/2026	Model	Depth Gauge
		PO#:		Gage	12"

Depth Gauge Readings As Found

Master	Actual	Deviation	Serial	Tol +0.001500 / -0.001500
+0.000000	+0.000000	+0.000000		
+0.500000	+0.500000	+0.000000		
+1.000000	+1.000000	+0.000000		
+6.000000	+5.999500	-0.000500		
+12.000000	+11.999500	-0.000500		

Comments

Procedure# SDM-49 Rev.1

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 5.3155HD16

Gage Block Set 46pc., Starrett Model Grade B

Traceable to NIST via NAMAS certification no: Cert#

CNB04-48449-711

Last / Next Cal Dates 10/31/2023 -> 10/31/2026

Gage Status PASS**Next Calibration Due: 01/08/2027**

Certified By: Mason Pellerin

Signature



This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self-calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees, Humidity 20-55%)

APPROVED**By Kayla Myers at 12:43 pm, Jul 10, 2026**