Gage ID: PEA-05-1556 Page 1 of 2

# TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

## **Certificate of Calibration**

### Certificate No TSP-07162025-009

Gage ID PEA-05-1556 Gage S/N 05-1556

**Description** 10-0-10 Gauss Meter

Operating OP-117 Procedure:

Unit of Meas. Gauss Manufacturer Magnaflux Cal. Date 7/16/2025

Next Due 1/16/2026

Cal. Freq. 6.00

**Months** 

Location Lab

**Environmental Conditions** 

Temperature 68 +/- 2 deg F

Humidity 20-55%

> Approved Yes Customer Info. Peak NDT

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### **Findings**

							~			
Standard ID	+5		Uncertainty			Units			Туре	٧
Limited Use?	No		Minimum		4	Nominal		5	Maximum	6
Ref Type	Gauss Meter		As Found		4	Accuracy		-1	Fail Before	No
			As Left		4	Accuracy		-1	Fail After	No
Gage ID of Star	ndard	Ref-09-1229								
Std Due Date		10/10/2025	Gage S/N	09-1229	NIST No.		114537			
Standard ID	+10		Uncertainty			Units			Туре	V
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11
Ref Type	Gauss	s Meter	As Found		9	Accuracy		-1	Fail Before	No
			As Left		9	Accuracy		-1	Fail After	No
Gage ID of Standard Ref-09-1229										
Std Due Date		10/10/2025	Gage S/N	09-1229	•	IIST No.	114537			
Standard ID	-5	<del></del>	Uncertainty			Units			Туре	V
Limited Use?	No		Minimum		4	Nominal		5	Maximum	6

Ref Type As Found 5 0 Fail Before No Gauss Meter Accuracy Fail After No As Left Accuracy Gage ID of Standard Ref-09-1229 Std Due Date 10/10/2025 Gage S/N 09-1229 NIST No. 114537 Standard ID -10 Uncertainty Units Type V Limited Use? No Minimum 9 Nominal 10 Maximum 11 Ref Type As Found 0 Gauss Meter 10 Accuracy Fail Before No As Left 10 Accuracy Fail After No Gage ID of Standard Ref-09-1229 Std Due Date 10/10/2025 Gage S/N 09-1229 NIST No. 114537

Page

Date:

2 of 2

7/16/2025

**APPROVED** 

Gage ID: PEA-05-1556

Calibrated By Taichi Daimo

By Kayla Myers at 4:32 pm, Jul 16, 2025

Signature