

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-07162025-010

Gage ID PEA-339814 Gage S/N 339814 Description 4 Step Block Operating Procedure: OP-126 Unit of Meas. inch Manufacturer N/A Cal. Date 7/16/2025 Next Due 7/16/2026 Cal. Freq. 1.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	.250 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.248	Nominal	0.250	Maximum	0.252
Ref Type	6" Digital Caliper	As Found	0.250	Accuracy	0.000	Fail Before	No
		As Left	0.250	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-08341639						
Std Due Date	10/10/2025	Gage S/N	08341639	NIST No.	CN304-48449-711		
Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	6" Digital Caliper	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-08341639						
Std Due Date	10/10/2025	Gage S/N	08341639	NIST No.	CN304-48449-711		
Standard ID	.750 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.748	Nominal	0.750	Maximum	0.752

Ref Type	6" Digital Caliper	As Found	0.750	Accuracy	0.000	Fail Before	No
		As Left	0.750	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/10/2025

Gage S/N 08341639

NIST No.

CN304-48449-711

Standard ID	1.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.998	Nominal	1.000	Maximum	1.002
Ref Type	6" Digital Caliper	As Found	1.000	Accuracy	0.000	Fail Before	No
		As Left	1.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/10/2025

Gage S/N 08341639

NIST No.

CN304-48449-711

Calibrated By Taichi Daimo

Signature

Date: 7/16/2025

APPROVED*By Kayla Myers at 4:33 pm, Jul 16, 2025*