

**TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

**Certificate of Calibration**

Certificate No TSP-03122025-005

<b>Gage ID</b> PEA-PCS31911 <b>Gage S/N</b> PCS31911 <b>Description</b> CD Box <b>Operating Procedure:</b> OP-120 <b>Unit of Meas.</b> Amps <b>Manufacturer</b> PCS <b>Cal. Date</b> 3/12/2025 <b>Next Due</b> 9/12/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 +/- 2 deg F <b>Humidity</b> 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
---	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

<b>Standard ID</b>	High 1st Shot	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	9360	<b>Nominal</b>	10400	<b>Maximum</b>	11440
<b>Ref Type</b>	Shot Check II	<b>As Found</b>	10600	<b>Accuracy</b>	200	<b>Fail Before</b>	No
		<b>As Left</b>	10600	<b>Accuracy</b>	200	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-SC-5100						
<b>Std Due Date</b>	10/22/2025	<b>Gage S/N</b>	5100	<b>NIST No.</b>	1112318 & 1112316		
<b>Standard ID</b>	High 2nd Shot	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	9360	<b>Nominal</b>	10400	<b>Maximum</b>	11440
<b>Ref Type</b>	Shot Check II	<b>As Found</b>	10600	<b>Accuracy</b>	200	<b>Fail Before</b>	No
		<b>As Left</b>	10600	<b>Accuracy</b>	200	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-SC-5100						
<b>Std Due Date</b>	10/22/2025	<b>Gage S/N</b>	5100	<b>NIST No.</b>	1112318 & 1112316		
<b>Standard ID</b>	High 3rd Shot	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	9360	<b>Nominal</b>	10400	<b>Maximum</b>	11440

Ref Type	Shot Check II	As Found	10600	Accuracy	200	Fail Before	No
		As Left	10600	Accuracy	200	Fail After	No
Gage ID of Standard	REF-SC-5100						
Std Due Date	10/22/2025	Gage S/N	5100	NIST No.	1112318 & 1112316		

Calibrated By Taichi Daimo Signature  Date: 3/12/2025

**APPROVED**  
*By Kayla Myers at 1:13 pm, Mar 12, 2025*