

SDM CALIBRATION\$ LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company Peak NDT Solutions
Address 515A Verot School Road
Lafayette LA
Zip 70506
Page 1 of 1

Doc #: 189884
OEM #:
Contact
Dept WO# 100626-13
Cal Date 10/12/2026
PO#:

Print#:
Serial#: 2104251917
Location Calibration Lab
Mfg.: Insize
Model Digital Caliper
Gage 12"

OD Jaw Values

Master	Actual	Deviation	Serial	Tol
+0.000000	+0.000000	+0.000000		+0.002000 / -0.002000
+0.500000	+0.500000	+0.000000		
+1.000000	+1.000000	+0.000000		
+6.000000	+6.000000	+0.000000		
+12.000000	+12.000500	+0.000500		

ID Jaw Value

Master	Actual	Deviation	Serial	Tol
+1.000000	+1.002000	+0.002000		+0.002000 / -0.002000

Comments

Procedure# SDM-44 Rev.1

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

Reference Standard Serial#: 5.3155HD16

Gage Block Set 46pc., Starrett Model Grade B

Traceable to NIST via NAMAS certification no: Cert#
CN304-48449-711

Last / Next Cal Dates: 10/31/2023 → 10/31/2026

Gage Status PASS

Next Calibration Due: 04/12/2027

Certified By: Mason Pellerin

Signature

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 2:41 pm, Oct 13, 2026