

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-05062026-000

<p>Gage ID PEA-11515-1 Gage S/N 11515-1 Description 20-0-20 Gauss Meter Operating Procedure: OP-117 Unit of Meas. Gauss Manufacturer R. B. Annis Co., Inc. Cal. Date 5/6/2026 Next Due 11/6/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID +10	Uncertainty	Units	Type V
Limited Use? No	Minimum 9	Nominal 10	Maximum 11
Ref Type Gauss Meter	As Found 10	Accuracy 0	Fail Before No
	As Left 10	Accuracy 0	Fail After No
Gage ID of Standard Ref-09-1229	Gage S/N 09-1229	NIST No. 114537	
Std Due Date 10/13/2026			

Standard ID +20	Uncertainty	Units	Type V
Limited Use? No	Minimum 19	Nominal 20	Maximum 21
Ref Type Gauss Meter	As Found 20	Accuracy 0	Fail Before No
	As Left 20	Accuracy 0	Fail After No
Gage ID of Standard Ref-09-1229	Gage S/N 09-1229	NIST No. 114537	
Std Due Date 10/13/2026			

Standard ID -10	Uncertainty	Units	Type V
Limited Use? No	Minimum 9	Nominal 10	Maximum 11

Ref Type	Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/13/2026

Gage S/N 09-1229

NIST No. 114537

Standard ID	-20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No

Gage ID of Standard Ref-09-1229

Std Due Date 10/13/2026

Gage S/N 09-1229

NIST No. 114537

Calibrated By Taichi Daimo Signature  Date: 5/6/2026

APPROVED
By Kayla Myers at 5:39 pm, May 07, 2026