

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-11102025-035

Gage ID PEA-47791 Gage S/N 47791 Description Pit Gauge Operating Procedure: OP-122 Unit of Meas. Inch Manufacturer Insize Cal. Date 11/10/2025 Next Due 11/10/2026 Cal. Freq. 1.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	.10000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.09900	Nominal	0.10000	Maximum	0.10100
Ref Type	Mic Trac	As Found	0.10008	Accuracy	0.00008	Fail Before	No
		As Left	0.10008	Accuracy	0.00008	Fail After	No
Gage ID of Standard	REF-G141308						
Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00		
Standard ID	.20000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.19900	Nominal	0.20000	Maximum	0.20100
Ref Type	Mic Trac	As Found	0.20015	Accuracy	0.00015	Fail Before	No
		As Left	0.20015	Accuracy	0.00015	Fail After	No
Gage ID of Standard	REF-G141308						
Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00		
Standard ID	.30000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.29900	Nominal	0.30000	Maximum	0.30100

Ref Type	Mic Trac	As Found	0.30014	Accuracy	0.00014	Fail Before	No
		As Left	0.30014	Accuracy	0.00014	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 10/1/2026

Gage S/N G141308

NIST No. 747V66-240502-00

Standard ID	.40000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.39900	Nominal	0.40000	Maximum	0.40100
Ref Type	Mic Trac	As Found	0.40017	Accuracy	0.00017	Fail Before	No
		As Left	0.40017	Accuracy	0.00017	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 10/1/2026

Gage S/N G141308

NIST No. 747V66-240502-00

Standard ID	.50000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.49900	Nominal	0.50000	Maximum	0.50100
Ref Type	Mic Trac	As Found	0.50007	Accuracy	0.00007	Fail Before	No
		As Left	0.50007	Accuracy	0.00007	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 10/1/2026

Gage S/N G141308

NIST No. 747V66-240502-00

Calibrated By	Taichi Daimo	Signature		Date:	11/10/2025
---------------	--------------	-----------	--	-------	------------

APPROVED*By Kayla Myers at 3:07 pm, Nov 12, 2025*