

**TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

**Certificate of Calibration**

Certificate No TSP-08112025-006

<b>Gage ID</b> PEA-11515-2 <b>Gage S/N</b> 11515-2 <b>Description</b> Profile Gauge <b>Operating Procedure:</b> OP-129 <b>Unit of Meas.</b> <b>Manufacturer</b> Lone star <b>Cal. Date</b> 8/11/2025 <b>Next Due</b> 8/11/2028 <b>Cal. Freq.</b> 3.00 Years <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 +/- 2 deg F <b>Humidity</b> 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
--	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

AS FOUND: PASSED AS LEFT: PASSED

Gage: NC-23-50, 4 TPI, 2 TPF

Profile Overlay Serial #: C14281  
Profile Overlay NIST# 03-7-1857

Optical Comparator w/ Digital Readout (14"), Mitutoyo, Model PH-A14  
Serial #: 012721312  
Traceable to NIST #: Cert# CO325-29789-775, CO325-28462-775  
Last / Next Cal Dates: 03/11/2025 --> 03/11/2028

Calibrated By Taichi Daimo

Signature

Date: 8/11/2025**APPROVED***By Kayla Myers at 11:41 am, Aug 11, 2025*