

**TSP Calibration, Inc.**

1110 S. FM 1788 Suite 38-40 Midland, TX 79706  
Phone: (432) 307-7054 Email: Chad@tspndt.com

**Certificate of Calibration**

Certificate No TSPW-10102023-002

<b>Gage ID</b> PEA-25431 <b>Gage S/N</b> 25431 <b>Description</b> 10 lb Test Weight <b>Operating Procedure:</b> OP-118 <b>Unit of Meas.</b> pounds <b>Manufacturer</b> Parker <b>Cal. Date</b> 10/10/2023 <b>Next Due</b> 10/10/2026 <b>Cal. Freq.</b> 3.00 Years <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 Deg F +/- 2 <b>Humidity</b> 35-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
--	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

SO 12162 Job 11300

AS FOUND: PASSED AS LEFT: PASSED

Using calibrated scale serial no. 8335029565

Test weight as found 10.075 lbs.

Test weight as left 10.075 lbs.

Calibration due date for scale is on 7/26/2024

Scale is traceable to NIST by the following numbers: Doc# 138436

Calibrated By Chad AdamsSignature Date: 10/10/2023**APPROVED****By Kayla Myers at 1:17 pm, Oct 17, 2023**