Tech Service Products

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland,TX 79706

Certificate of Calibration

Certificate No TSPW-10262026-004

Gage ID	PEA-CES-105661
Gage S/N	CES-105661
Description	10-1/2" Magnetic Coil
Operating Procedure:	OP-116
Unit of Meas.	Gauss
Manufacturer	Cut-N-Edge
Cal. Date	10/26/2026
Next Due	4/26/2027
Cal. Freq.	6.00 Months
Location	Lab

Environmental Conditions

Temperature	68 Deg F +/- 2
Humidity	20-55%

Approved	Yes
Customer Info.	Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No. 1067233 Job No. 15309

As Found: Passed As Left: Passed

Coil putting out magnetic field from both the positive and negative side equal to or greater than 50 gauss for calibration.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
Reference Standard: 50-0-50 Magnetometer - Serial Number: TSPSTD-01
NIST Number: 187828 - Calibration Due Date: 8/27/2027

Calibrated By **Chad Adams** **Signature**  **Date:** **10/26/2026**

APPROVED
By Kayla Myers at 11:23 am, Oct 29, 2026