Tech Service Products

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland,TX 79706

Certificate of Calibration

Certificate No TSPW-10262026-005

Gage ID	PEA-N7362
Gage S/N	N7362
Description	A/C Yoke
Operating Procedure:	OP-115
Unit of Meas.	Pounds
Manufacturer	Magnaflux
Cal. Date	10/26/2026
Next Due	4/26/2027
Cal. Freq.	6.00 Months
Location	Lab

Environmental Conditions

Temperature	68 Deg F +/- 2
Humidity	20-55%

Approved	Yes
Customer Info.	Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.


Findings

Order No. 1067233 Job No. 15309

As Found: Passed As Left: Passed

A/C Yoke picked up and held required 10 pound weight for calibration.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
Reference Standard: Test Weight - Serial Number: 3452
NIST Number: Doc#171848 - Calibration Due Date: 8/21/2027

Calibrated By **Chad Adams** **Signature**  **Date:** **10/26/2026**

APPROVED
By Kayla Myers at 11:28 am, Oct 29, 2026