

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-06172025-000

<b>Gage ID</b> PEA-13005 <b>Gage S/N</b> 13005 <b>Description</b> 50-0-50 Gauss Meter <b>Operating Procedure:</b> OP-117 <b>Unit of Meas.</b> Gauss <b>Manufacturer</b> R. B. Annis Co., Inc. <b>Cal. Date</b> 6/17/2025 <b>Next Due</b> 12/17/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 +/- 2 deg F <b>Humidity</b> 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
--	---

#### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### Findings

<b>Standard ID</b>	+25	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	24	<b>Nominal</b>	25	<b>Maximum</b>	26
<b>Ref Type</b>	Gauss Meter	<b>As Found</b>	25	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	25	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-09-1234						
<b>Std Due Date</b>	5/12/2026	<b>Gage S/N</b>	09-1234	<b>NIST No.</b>	1114538		
<b>Standard ID</b>	+50	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	49	<b>Nominal</b>	50	<b>Maximum</b>	51
<b>Ref Type</b>	Gauss Meter	<b>As Found</b>	50	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	50	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-09-1234						
<b>Std Due Date</b>	5/12/2026	<b>Gage S/N</b>	09-1234	<b>NIST No.</b>	1114538		
<b>Standard ID</b>	-25	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	24	<b>Nominal</b>	25	<b>Maximum</b>	26

Gage ID: PEA-13005

Page 2 of 2

Ref Type	Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No

Gage ID of Standard REF-09-1234

Std Due Date	5/12/2026	Gage S/N	09-1234	NIST No.	1114538
--------------	-----------	----------	---------	----------	---------

Standard ID	-50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No

Gage ID of Standard REF-09-1234

Std Due Date	5/12/2026	Gage S/N	09-1234	NIST No.	1114538
--------------	-----------	----------	---------	----------	---------

Calibrated By	Taichi Daimo	Signature		Date:	6/17/2025
---------------	--------------	-----------	--	-------	-----------

**APPROVED**

*By Kayla Myers at 12:25 pm, Jun 17, 2025*