

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-04102025-005

Gage ID PEA-21169 Gage S/N 21169 Description 10 lb. Test Weight Operating Procedure: OP-118 Unit of Meas. lbs Manufacturer Parker Cal. Date 4/10/2025 Next Due 4/10/2028 Cal. Freq. 3.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

AS FOUND: PASSED AS LEFT: PASSED

Using calibrated scale serial no. 7129341906

Test weight as found 10.07 lbs.

Test weight as left 10.07 lbs.

Calibration due date for scale is on 4/15/2025.

Scale is traceable to NIST by the following numbers: LA22-258

Calibrated By Taichi Daimo

Signature

Date: 4/10/2025**APPROVED***By Kayla Myers at 12:26 pm, Apr 10, 2025*