Tubular Inspection Products
1551 Windway
Odessa, TX 79763
(432) 363-2200
PO Box 862
Odessa, TX 79760
Fax: (432) 363-2202

CERTIFICATE OF CALIBRATION

Traceable to NIST — ISO/IEC 17025:2017

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO 17025:2017 • ISO 9001 • API Q1/7-2

INSTRUMENT & CALIBRATION DETAILS

Customer	Peak NDT Solutions, LLC	**Manufacturer**	General
Calibration Date	July 21, 2026	**TIP Procedure**	TIP-UVMTW01.1
Calibration Due Date	January 21, 2027	**Calibration Technician**	Ryan Britton
Serial Number	2225000429	**Order Number**	81039
Instrument	UV513AB UV Light Meter	**Temperature / Humidity**	68.5°F / RH 40.5%

CALIBRATION RESULTS

Distance	Calibrated Reading	Instrument Under Test	Tolerance / Result
15"	9375 µw/cm²	9369 µw/cm²	**±5% — PASS**

INSTRUMENT STATUS SUMMARY

Instrument Received	Instrument Returned	Instrument Status
In Tolerance	In Tolerance	**PASSED**

CALIBRATION STANDARDS USED (TRACEABILITY)

- TIP Master Blacklight, Spectroline Tritan 365MHB, S/N 1812802 (Cert. Due 01/28/2027)
- Spectroline XDS-1000 UV, S/N 1943686 (Cert. Due 01/28/2027)

Calibration Technician

Ryan Britton

Issue Date: July 21, 2026

APPROVED
By Kayla Myers at 3:13 pm, Nov 09, 2026