

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration**Certificate No TSP-06162025-019**

Gage ID PEA-2225000429 Gage S/N 2225000429 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 6/16/2025 Next Due 12/16/2025 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Calibrated black light serial no. 2234240 was also used during this calibration.
Calibration due date for black light is on 12/13/2025

Black light is traceable to NIST by the following numbers: 258648, 258649

Standard ID 12"	Uncertainty	Units	Type V
Limited Use? No	Minimum 6466	Nominal 6806	Maximum 7146
Ref Type Long Wave	As Found 6938	Accuracy 132	Fail Before No
	As Left 6938	Accuracy 132	Fail After No
Gage ID of Standard REF-1944278			
Std Due Date 6/17/2025	Gage S/N 1944278/1944 280	NIST No. 258648, 258649, 258650	

Standard ID 15"	Uncertainty	Units	Type V
Limited Use? No	Minimum 4349	Nominal 4577	Maximum 4805
Ref Type Long Wave	As Found 4708	Accuracy 131	Fail Before No
	As Left 4708	Accuracy 131	Fail After No
Gage ID of Standard REF-1944278			
Std Due Date 6/17/2025	Gage S/N 1944278/1944 280	NIST No. 258648, 258649, 258650	

Calibrated By Taichi Daimo

Signature



Date: 6/16/2025

APPROVED

By Kayla Myers at 12:23 pm, Jun 17, 2025