



Tubular Inspection Products Inc.

Certificate of Calibration

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO17025:2017/ISO9001/API Q1/7-2

Customer: **Peak NDT Solutions**

Calibration Date: **July 14, 2026**

Calibration Due Date: **January 14, 2027**

Serial Number: **N5982**

Instrument: **Y-2 AC Yoke**

Manufacturer: **Magnaflux**

TIP Procedure: **TIP-YKW01.1**

Calibration Technician: **Ryan Britton**

Order Number: **80910**

Results

<u>Lift Test Reference</u>	<u>Actual Measurement</u>	<u>Deviation</u>
Lift > 3sec.	Lift > 3 sec.	N/A
<u>Tolerance</u>		<u>Must Lift pass 3 second</u>

<u>Instrument Received</u>	<u>Instrument Returned</u>	<u>Instrument Status</u>
In Tolerance	In Tolerance	Passed

Temperature: 68.8°F, RH 38.5%

Issue Date for Certificate: **July 14, 2026**

TIP Master Optical Comparator, Keyence IM-8030

SN: 4C0Q000412 Cal Date: 06-26-2026

TIP -MS-10TW SN: 33385 Cal Date: 05-05-2026

TIP Master Hygrometer SN: 31874/1H2048MCR, 130070751, Cal Date: 09-09-2026

Approved By: _____

APPROVED

By Kayla Myers at 8:55 am, Jul 24, 2026

Tubular Inspection Products
1551 Windway St. Odessa, Tx 79763
Calibrations: Andres Ramirez O. 432-363-2200 C. 432-924-3532
Email: Andy@tubularip.com