

# TSP Calibration, Inc.

1110 S. FM 1788 #42  
Phone: (432) 296-4992 Email: Chad@tspndt.com  
Midland, TX 79706

## Certificate of Calibration

Certificate No TSPW-01302026-015

<p><b>Gage ID</b> PEA-N5982  <b>Gage S/N</b> N5982  <b>Description</b> A/C Yoke  <b>Operating Procedure:</b> OP-115  <b>Unit of Meas.</b> pounds  <b>Manufacturer</b> Magnaflux  <b>Cal. Date</b> 1/30/2026  <b>Next Due</b> 7/30/2026  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>  Temperature 68 Deg F +/- 2  Humidity 35-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
--	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings


Job 15136

As Found: Passed As Left: Passed

A/C Yoke picked up and held required 10 pound weight for calibration.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: Test Weight - Serial Number: 3452  
NIST Number: Doc#171848 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams Signature  Date: 1/30/2026

**APPROVED**  
By Kayla Myers at 4:00 pm, Feb 06, 2026