



Tubular Inspection Products Inc.

Certificate of Calibration

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO17025:2017/ISO9001/API Q1/7-2

Customer: **Peak NDT Solutions**

Calibration Date: July 13, 2026

Calibration Due Date: January 13, 2027

Serial Number: 1402232236

Instrument: 12" Digital Caliper

Manufacturer: Insize

TIP Procedure: TIP-CALW01.1

Calibration Technician: Ryan Britton

Order Number: 80910

Results

	Reference Size	Actual Measurement	Deviation	Tolerance	Result
Step	<u>1.000"</u>	<u>1.0005"</u>	<u>0.0005"</u>	<u>+/-0.001"</u>	<u>Pass</u>
Step	<u>2.000"</u>	<u>2.0000"</u>	<u>0.0000"</u>	<u>+/-0.001"</u>	<u>Pass</u>
Step	<u>3.000"</u>	<u>3.0005"</u>	<u>0.0005"</u>	<u>+/-0.001"</u>	<u>Pass</u>
Step	<u>4.000"</u>	<u>4.0005"</u>	<u>0.0005"</u>	<u>+/-0.001"</u>	<u>Pass</u>

<u>Instrument Received</u>	<u>Instrument Returned</u>	<u>Instrument Status</u>
In Tolerance	In Tolerance	Passed

Temperature/Humidity: 68.5°F / RH 40.5%

Issue Date for Certificate: July 13, 2026

TIP Master Optical Comparator, Keyence IM-8030,

4C0Q000412 Cal Date: 06-26-2027

TIP Masters Blocks : 13232,14002,14340,13207 Cal Date: 01-25-2027

TIP Master Hygrometer SN: 31874/1H2048MCR, 130070751, Cal Date: 01-29-27

Approved By: _____

APPROVED

By Kayla Myers at 8:53 am, Jul 24, 2026