# Tech Service Products

**1110 S. FM 1788 #42**
**Phone: (432) 296-4992 Email: Chad@tspndt.com**
**Midland,TX 79706**

## Certificate of Calibration

**Certificate No TSPW-10132026-006**

| | |
|---|---|
| **Gage ID** | PEA-12679-1 |
| **Gage S/N** | 12679-1 |
| **Description** | 20-0-20 Gauss Meter |
| **Operating Procedure:** | OP-117 |
| **Unit of Meas.** | Gauss |
| **Manufacturer** | R.B. Annis |
| **Cal. Date** | 10/13/2026 |
| **Next Due** | 4/13/2027 |
| **Cal. Freq.** | 6.00 Months |
| **Location** | Lab |

**Environmental Conditions**

| | |
|---|---|
| **Temperature** | 68 Deg F +/- 2 |
| **Humidity** | 20-55% |

**Approved** Yes
**Customer Info.** Peak NDT Solutions

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

Order No. 1066378 Job No. 15301

| | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| **Standard ID** | +10 | **Uncertainty** | | | **Units** | | | **Type** | V |
| **Limited Use?** | No | **Minimum** | 9 | | **Nominal** | 10 | | **Maximum** | 11 |
| **Ref Type** | 20-0-20 Gauss Meter | **As Found** | 10 | | **Accuracy** | 0 | | **Fail Before** | No |
| | | **As Left** | 10 | | **Accuracy** | 0 | | **Fail After** | No |
| **Gage ID of Standard** | REF-TSPSTD-02 | | | | | | | | |
| **Std Due Date** | 8/27/2027 | **Gage S/N** | TSPSTD-02 | | **NIST No.** | 187829 | | | |

| | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| **Standard ID** | +20 | **Uncertainty** | | | **Units** | | | **Type** | V |
| **Limited Use?** | No | **Minimum** | 19 | | **Nominal** | 20 | | **Maximum** | 21 |
| **Ref Type** | 20-0-20 Gauss Meter | **As Found** | 20 | | **Accuracy** | 0 | | **Fail Before** | No |
| | | **As Left** | 20 | | **Accuracy** | 0 | | **Fail After** | No |
| **Gage ID of Standard** | REF-TSPSTD-02 | | | | | | | | |
| **Std Due Date** | 8/27/2027 | **Gage S/N** | TSPSTD-02 | | **NIST No.** | 187829 | | | |

| | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|
| **Standard ID** | -10 | **Uncertainty** | | | **Units** | | | **Type** | V |
| **Limited Use?** | No | **Minimum** | 9 | | **Nominal** | 10 | | **Maximum** | 11 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Ref Type** | 20-0-20 Gauss Meter | **As Found** | 10 | **Accuracy** | 0 | **Fail Before** | No |
| | | **As Left** | 10 | **Accuracy** | 0 | **Fail After** | No |

| | | | | | |
|---|---|---|---|---|---|
| **Gage ID of Standard** | REF-TSPSTD-02 | | | | |
| **Std Due Date** | 8/27/2027 | **Gage S/N** | TSPSTD-02 | **NIST No.** | 187829 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **Standard ID** | -20 | **Uncertainty** | | **Units** | | **Type** | V |
| **Limited Use?** | No | **Minimum** | 19 | **Nominal** | 20 | **Maximum** | 21 |
| **Ref Type** | 20-0-20 Gauss Meter | **As Found** | 20 | **Accuracy** | 0 | **Fail Before** | No |
| | | **As Left** | 20 | **Accuracy** | 0 | **Fail After** | No |

| | | | | | |
|---|---|---|---|---|---|
| **Gage ID of Standard** | REF-TSPSTD-02 | | | | |
| **Std Due Date** | 8/27/2027 | **Gage S/N** | TSPSTD-02 | **NIST No.** | 187829 |

**Calibrated By** Chad Adams **Signature** _____ **Date:** 10/13/2026

***APPROVED***
***By Kayla Myers at 8:48 am, Oct 14, 2026***