

## TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706  
 Phone: (432) 296-4992 Email: Chad@tspndt.com

## Certificate of Calibration

Certificate No TSPW-01232025-017

<b>Gage ID</b> PEA-12679-1 <b>Gage S/N</b> 12679-1 <b>Description</b> 20-0-20 gauss meter <b>Operating Procedure:</b> OP-117 <b>Unit of Meas.</b> gauss <b>Manufacturer</b> R.B. Annis <b>Cal. Date</b> 1/23/2025 <b>Next Due</b> 7/23/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 Deg F +/- 2 <b>Humidity</b> 35-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT Solutions
--	---

## Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

## Findings

SO 13802 Job 11764

Standard ID	+10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11
Ref Type	20-0-20 Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No
Gage ID of Standard	REF-TSPSTD-02						
Std Due Date	8/8/2025	Gage S/N	TSPSTD-02	NIST No.	154784		

Standard ID	+20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	20-0-20 Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No
Gage ID of Standard	REF-TSPSTD-02						
Std Due Date	8/8/2025	Gage S/N	TSPSTD-02	NIST No.	154784		

Standard ID	-10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11

Ref Type	20-0-20 Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No

Gage ID of Standard REF-TSPSTD-02

Std Due Date 8/8/2025

Gage S/N TSPSTD-02 NIST No. 154784

Standard ID	-20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	20-0-20 Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No


Gage ID of Standard REF-TSPSTD-02

Std Due Date 8/8/2025

Gage S/N TSPSTD-02 NIST No. 154784

Calibrated By Chad Adams

Signature



Date: 1/23/2025

**APPROVED**

By Kayla Myers at 8:20 am, Apr 14, 2025