

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-02232026-001

<p>Gage ID PEA-7632 Gage S/N 7632 Description 20-0-20 gauss meter Operating Procedure: OP-117 Unit of Meas. gauss Manufacturer R.B. Annis Cal. Date 2/23/2026 Next Due 8/23/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15584 Job 15174

Standard ID	+10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11
Ref Type	20-0-20 Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No
Gage ID of Standard	REF-TSPSTD-02						
Std Due Date	8/21/2026	Gage S/N	TSPSTD-02	NIST No.	171844		

Standard ID	+20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	20-0-20 Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No
Gage ID of Standard	REF-TSPSTD-02						
Std Due Date	8/21/2026	Gage S/N	TSPSTD-02	NIST No.	171844		

Standard ID	-10	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	9	Nominal	10	Maximum	11

Ref Type	20-0-20 Gauss Meter	As Found	10	Accuracy	0	Fail Before	No
		As Left	10	Accuracy	0	Fail After	No

Gage ID of Standard REF-TSPSTD-02

Std Due Date	8/21/2026	Gage S/N	TSPSTD-02	NIST No.	171844
--------------	-----------	----------	-----------	----------	--------

Standard ID	-20	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	19	Nominal	20	Maximum	21
Ref Type	20-0-20 Gauss Meter	As Found	20	Accuracy	0	Fail Before	No
		As Left	20	Accuracy	0	Fail After	No

Gage ID of Standard REF-TSPSTD-02

Std Due Date	8/21/2026	Gage S/N	TSPSTD-02	NIST No.	171844
--------------	-----------	----------	-----------	----------	--------

Calibrated By Chad Adams Signature  Date: 2/23/2026

APPROVED
By Kayla Myers at 10:01 pm, Jun 02, 2026