Gage ID: PEA-12679-1

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706 Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-08252025-007

Gage ID PEA-12679-1

Gage S/N 12679-1

Description 20-0-20 gauss meter

Operating OP-117

Procedure:

Unit of Meas. gauss

Manufacturer R.B. Annis

Cal. Date 8/25/2025 Next Due 2/25/2026

Cal. Freq. 6.00

Months

Location Lab

Environmental Conditions

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14759 Job 14274

- L. L.D.	.10		Uncertainty			Units			Type	V	
Standard ID	+10		Minimum		9	Nominal		10	Maximum	11	
Limited Use?	No 0.3	0 Gauss Meter	As Found		10	Accuracy		0	Fail Before	No	
Ref Type	20-0-2	o Gauss Meter	As Left		10	Accuracy		0	Fail After	No	
Gage ID of Standard		REF-TSPSTD-02					474044				
Std Due Date		8/21/2026	Gage S/N	TSPSTD-02	1	NIST No.	171844				
	+20		Uncertainty			Units			Туре	٧	
Standard ID			Minimum		19	Nominal		20	Maximum	21	
Limited Use?	No	20 Gauss Meter	As Found		20	Accuracy		0	Fail Before	No	
Ref Type 20-0		-20 Gauss Meter	As Left		20	Accuracy		0	Fail After	No	
Gage ID of Standard REF-TSPSTD		REF-TSPSTD-02									
Std Due Date		8/21/2026	Gage S/N	TSPSTD-02	NIST No.		171844				
									Type	V	
Standard ID	-10		Uncertainty			Units		10	Maximum		
Limited Use?	No		Minimum		9	Nominal		10	IVIAXIMUM	- ' '	

Fail Before No Accuracy As Found 10 20-0-20 Gauss Meter Ref Type Fail After No Accuracy 10 As Left REF-TSPSTD-02 Gage ID of Standard NIST No. 171844 Gage S/N TSPSTD-02 8/21/2026 Std Due Date Type V Units Uncertainty -20 Standard ID Maximum 21 20 Nominal Minimum 19 Limited Use? No 0 Fail Before No Accuracy 20 As Found 20-0-20 Gauss Meter Ref Type Fail After No 0 As Left 20 Accuracy REF-TSPSTD-02 Gage ID of Standard 171844 NIST No. Gage S/N TSPSTD-02 8/21/2026 Std Due Date

Calibrated By Chad Adams

Gage ID: PEA-12679-1

Signature

Date:

Page

8/25/2025

2 of 2

APPROVED

By Kayla Myers at 10:48 am, Aug 26, 2025