Tech Service Products

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland,TX 79706

Certificate of Calibration

Certificate No TSPW-10132026-013

Gage ID	PEA-230721047
Gage S/N	230721047
Description	0-1" Groove Micrometer
Operating Procedure:	OP-110
Unit of Meas.	Inch
Manufacturer	Insize
Cal. Date	10/13/2026
Next Due	4/13/2027
Cal. Freq.	6.00 Months
Location	Lab

Environmental Conditions

Temperature	68 Deg F +/- 2
Humidity	20-55%

Approved Yes
Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No. 1066378 Job No. 15301

Standard ID	0.100"	**Uncertainty**		**Units**		**Type**	V	
Limited Use?	No	**Minimum**	0.099	**Nominal**	0.100	**Maximum**	0.101	
Ref Type	81PC Gage Blocks	**As Found**	0.100	**Accuracy**	0.000	**Fail Before**	No	
		As Left	0.100	**Accuracy**	0.000	**Fail After**	No	
Gage ID of Standard	REF-070111							
Std Due Date	1/26/2027	**Gage S/N**	070111	**NIST No.**	81004			

Standard ID	0.300"	**Uncertainty**		**Units**		**Type**	V	
Limited Use?	No	**Minimum**	0.299	**Nominal**	0.300	**Maximum**	0.301	
Ref Type	81PC Gage Blocks	**As Found**	0.300	**Accuracy**	0.000	**Fail Before**	No	
		As Left	0.300	**Accuracy**	0.000	**Fail After**	No	
Gage ID of Standard	REF-070111							
Std Due Date	1/26/2027	**Gage S/N**	070111	**NIST No.**	81004			

Standard ID	0.500"	**Uncertainty**		**Units**		**Type**	V	
Limited Use?	No	**Minimum**	0.499	**Nominal**	0.500	**Maximum**	0.501	

Ref Type	81PC Gage Blocks	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No


Gage ID of Standard REF-070111
Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004

Standard ID	0.700"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.699	Nominal	0.700	Maximum	0.701
Ref Type	81PC Gage Blocks	As Found	0.700	Accuracy	0.000	Fail Before	No
		As Left	0.700	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111
Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004

Standard ID	0.900"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.899	Nominal	0.900	Maximum	0.901
Ref Type	81PC Gage Blocks	As Found	0.900	Accuracy	0.000	Fail Before	No
		As Left	0.900	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111
Std Due Date 1/26/2027 Gage S/N 070111 NIST No. 81004

Calibrated By Chad Adams Signature  Date: 10/13/2026

APPROVED
By Kayla Myers at 8:51 am, Oct 14, 2026